

# Electrical and Ultrasonic Bone Growth Stimulators (for Kansas Only)

**Policy Number:** CS037KS.02  
**Effective Date:** February 1, 2026

[Instructions for Use](#)

Table of Contents	Page
<a href="#">Application</a> .....	1
<a href="#">Coverage Rationale</a> .....	1
<a href="#">Medical Records Documentation Used for Reviews</a> .....	2
<a href="#">Definitions</a> .....	2
<a href="#">Applicable Codes</a> .....	2
<a href="#">Description of Services</a> .....	3
<a href="#">Clinical Evidence</a> .....	3
<a href="#">U.S. Food and Drug Administration</a> .....	5
<a href="#">References</a> .....	5
<a href="#">Policy History/Revision Information</a> .....	5
<a href="#">Instructions for Use</a> .....	6

Related Policies
None

## Application

This Medical Policy only applies to the state of Kansas.

## Coverage Rationale

### Spinal Applications

#### *Invasive Electrical*

The use of an **Invasive** spinal **Electrical Bone Growth Stimulator** is proven and medically necessary as an adjunct to lumbar spinal fusion surgery when the following criteria are met:

- Radiographic evidence of skeletal maturity; and
- Increased risk for fusion failure demonstrated by any of the following:
  - Previously failed fusion at the same site, when minimum of six months has elapsed since the last surgical procedure
  - Spinal fusion performed or to be performed at more than one level as part of a single surgery
  - Comorbid conditions associated with compromised bone healing (e.g., diabetes, obesity, osteoporosis, current tobacco use)
  - Spondylolisthesis grade II or greater

#### *Noninvasive Electrical*

The use of noninvasive spinal **Electrical Bone Growth Stimulator** is proven and medically necessary as an adjunct to spinal surgery under certain circumstances. For medical necessity clinical coverage criteria, refer to the InterQual® CP: Durable Medical Equipment, Bone Growth Stimulators, Noninvasive.

[Click here to view the InterQual® criteria.](#)

### Non-Spinal Applications, Invasive, and Noninvasive

The use of **Invasive** or **noninvasive** bone growth stimulators for non-spinal applications is considered proven and medically necessary in certain circumstances. For medical necessity clinical coverage criteria, refer to the InterQual® CP:

- Durable Medical Equipment, Bone Growth Stimulators, Noninvasive
- Procedures, Bone Graft and Implantable Stimulator, Fracture Nonunion

[Click here to view the InterQual® criteria.](#)

## Medical Records Documentation Used for Reviews

Benefit coverage for health services is determined by the federal, state, or contractual requirements, and applicable laws that may require coverage for a specific service. Medical records documentation may be required to assess whether the member meets the clinical criteria for coverage but does not guarantee coverage of the services requested.

The patient's medical record must contain documentation that fully supports the medical necessity for the requested services. This documentation includes, but is not limited to, relevant medical history, physical examination, and results of pertinent diagnostic tests or procedures. Documentation supporting the medical necessity should be legible, maintained in the patient's medical record, and must be made available upon request.

## Definitions

**Electrical Bone Growth Stimulator:** A device (either implanted into the body or worn externally) that uses an electric field or current to stimulate the growth of bone tissue (Haglin 2017).

**Invasive:** The implantable current generator is surgically placed in an intramuscular or subcutaneous space, while an electrode is implanted within the fragments of bone graft at the targeted fusion site. The implanted device is usually functional for 6 to 9 months at which point the current generator is removed in a second surgical procedure, while the electrodes may or may not be removed. Implantable bone growth stimulators are used as an adjunct to spinal fusion surgery and implanted at the time of surgery.

## Applicable Codes

The following list(s) of procedure and/or diagnosis codes is provided for reference purposes only and may not be all inclusive. Listing of a code in this policy does not imply that the service described by the code is a covered or non-covered health service. Benefit coverage for health services is determined by federal, state, or contractual requirements and applicable laws that may require coverage for a specific service. The inclusion of a code does not imply any right to reimbursement or guarantee claim payment. Other Policies and Guidelines may apply.

CPT Code	Description
<b>Electrical Bone Growth Stimulator: Non-Spinal (Invasive, Non-Invasive)</b>	
20974	Electrical stimulation to aid bone healing; noninvasive (nonoperative)
20975	Electrical stimulation to aid bone healing; invasive (operative)
<b>Ultrasound Bone Growth Stimulator</b>	
20979	Low intensity ultrasound stimulation to aid bone healing, noninvasive (nonoperative)

*CPT® is a registered trademark of the American Medical Association*

**Coding Clarification:** Utilize HCPCS code E0748 when reporting bone growth stimulation for all anatomical levels of the spine.

HCPCS Code	Description
E0747	Osteogenesis stimulator, electrical, noninvasive, other than spinal applications
E0748	Osteogenesis stimulator, electrical, noninvasive, spinal applications
E0749	Osteogenesis stimulator, electrical, surgically implanted
E0760	Osteogenesis stimulator, low intensity ultrasound, noninvasive

## Description of Services

Bone healing is a complex process dependent on a variety of factors. The rate of bone repair and composition of tissue varies depending on type of bone fractured, the extent of the bone and soft tissue damage, the adequacy of the blood supply, and the degree of separation between bone ends. The individual's general health and nutritional status also play a significant role in bone healing. The presence of infection may adversely affect healing. Diminished blood flow to the fracture site will often suppress the healing response. Factors that can cause diminished blood flow include heavy smoking, malnutrition, diabetes, alcoholism, peripheral vascular disease, age, and the use of some medications such as steroids. Other characteristics such as high-grade trauma, high grade and open fractures, comminution of the fracture, vertical or oblique fracture pattern, and fracture displacement may also contribute to poor healing of bone [Agency for Healthcare Research and Quality (AHRQ), 2005].

Bone growth stimulation is utilized to promote bone healing in difficult to heal fractures or fusions by applying electrical or ultrasonic current to the fracture/fusion site. Ultrasonic stimulation is applied externally, while electrical stimulation can be applied either from the outside of the body (noninvasive) or from the inside of the body (Invasive). (ECRI, 2024).

Bone growth stimulators are only indicated for use in individuals who are skeletally mature. A person is said to be skeletally mature when all bone growth is complete; the cartilage cells of the growth plate cease to proliferate, the growth plate becomes thinner, is replaced by bone and disappears, and the epiphysis is "closed" or fused with the shaft. [Agency for Healthcare Research and Quality (AHRQ), 2005].

## Clinical Evidence

### Electrical Bone-Growth Stimulators (EBGS)

Akhter et al. (2020) conducted a systematic review and meta-analysis to evaluate if postoperative electrical stimulation is more efficacious than no stimulation or placebo in promoting radiographic fusion in patients undergoing spinal fusion. The investigators searched the Cochrane Central Register of Controlled Trials (CENTRAL), EMBASE, CINAHL, and MEDLINE from date of inception to current. Ongoing clinical trials were also identified, and reference lists of included studies were manually searched for relevant articles. Two reviewers independently screened studies, extracted data, and assessed risk of bias. Data were pooled using the Mantel-Haenszel method. Trialists were contacted for any missing or incomplete data. Of 1,184 articles screened, seven studies were eligible for final inclusion (n = 941). A total of 487 patients received postoperative electrical stimulation and 454 patients received control or sham stimulation. All evidence was of moderate quality. Electrical stimulation (pulsed electromagnetic fields, direct current, and capacitive coupling) increased the odds of a successful fusion by 2.5-fold relative to control. A test for subgroup interaction by stimulation type, smoking status, and number of levels fused was not significant. The investigators concluded that this systematic review and meta-analysis found moderate-level evidence supporting the use of postoperative electrical stimulation as an adjunct to spinal fusion surgery. When compared to sham, placebo-controlled, or no stimulation, patients treated with postoperative electrical stimulation have significantly greater rates of successful radiographically defined fusions. According to the investigators, these results are supported by a notably high statistically significant effect, a narrow confidence interval, and the inclusion of only high-quality randomized trials with human subjects.

Aleem et al. (2016) conducted a meta-analysis of randomized sham-controlled trials to assess the efficacy of electrical stimulation for bone healing. Outcomes were pain relief, functional improvement, and radiographic nonunion. Fifteen trials met the inclusion criteria. Four trials included patients undergoing spinal fusion, five trials evaluated fresh fracture treatment, five trials examined treatment of delayed or nonunion, and one study included patients undergoing surgical osteotomy. According to the investigators, this systematic review and meta-analysis showed that patients treated with electrical stimulation as an adjunct for bone healing experience lower rates of radiographic nonunion or persistent nonunion and have significantly less pain. The results related to pain reduction were, however, mainly due to one trial that evaluated electrical stimulation for spinal fusion surgery. The trials that evaluated electrical stimulation for nonunion or delayed union of long bones did not show statistically significant pain reduction nor a statistically significant difference at the level of meta-analysis between treatment and control groups for the radiographic union outcome (risk ratio for persistent nonunion compared to sham: 0.55, 95%CI: 0.29 to 1.12). The investigators also indicated that no difference was seen in functional outcomes in a limited number of trials. According to the investigators, future trials focusing on functional outcomes are needed.

A systematic review of electrical stimulation to enhance bone healing by Griffin and Bayat identified 105 clinical studies and 35 in vitro studies of the technology. Direct current was found to be effective in enhancing bone healing in spinal fusion, as supported by four studies at level of evidence 1 (randomized control trial). The authors found support for its use for nonunion fractures, but only based on level of evidence 4 (case series). Eleven studies were retrieved for capacitive

coupling suggesting its effectiveness for spinal fusion but, for treating nonunions, the findings were conflicting. Studies of inductive coupling for long bones had conflicting findings. Overall, the studies, although in favor of electrical stimulation application in bone repair, displayed variability in treatment regime, primary outcome measures, follow-up times, and study design, making critical evaluation and assessment difficult (Griffin and Bayat, 2011).

A randomized controlled trial by Foley et al. (2008) tested the efficacy of PEMF stimulation to support cervical fusion in 323 participants with compressed cervical nerve root and symptomatic radiculopathy appropriate to the compressed root that had failed to respond to nonoperative management. While the group randomized to PEMF showed a significantly higher fusion rate than the control group (83.6% vs. 68.6%,  $p = 0.0065$ ) at six months, the group difference disappeared at 12 months post-surgery (92.8% vs. 86.7%,  $p = 0.1129$ ). Additionally, the study failed to show any group difference in patient-centered outcome such as pain scores, neck disability index, or functional status at six or 12 months. The authors concluded that although PEMF stimulation appeared to hasten bone healing in this randomized trial, it did not result in a significant advantage in terms of ultimate fusion rates or clinical outcomes for cervical fusion.

## ***Clinical Practice Guidelines***

### **North American Spine Society (NASS)**

In an updated 2025 coverage policy recommendation, NASS makes the following recommendations for electrical stimulation for bone healing:

- Electrical stimulation for spinal fusion healing is indicated for the following with qualifying criteria, when appropriate:
  - For augmentation of spinal fusion in any and all regions of the spine including occipital-cervical, cervical, cervicothoracic, thoracic, thoracolumbar, lumbar, and lumbosacral spinal regions in patients at high-risk for the development of pseudarthrosis (i.e., nonunion) who exhibit one or more of the following:
    - Are undergoing spinal fusion of 2 or more motion segments (3 vertebrae)
    - Are undergoing a revision spinal fusion (e.g., repeat surgery for a previously unhealed fusion attempt)
    - Are smokers who cannot stop smoking in preparation for fusion due to the nature of the underlying condition (e.g., acute traumatic fracture)
    - Exhibit one or more of the following comorbidities when undergoing primary lumbar fusion:
      - Diabetes
      - Inflammatory arthritis (e.g., rheumatoid arthritis) that has required long-term corticosteroid therapy
      - Immunocompromised (e.g., undergoing chemotherapy and radiation therapy to the spine, hypogammaglobulinemia, granulocytopenia, acquired immune deficiency syndrome, chronic granulomatous disease)
      - Systemic vascular disease
      - Osteopenia or osteoporosis
  - Electrical stimulation of spinal fusion is not indicated in the following clinical scenarios:
    - Primary (index or first-attempt) spinal fusions without additional risk factors listed above
    - Spinal fusion of 1 motion segment or 2 vertebrae levels without additional risk factors listed above
    - Presence of malignancy
    - Presence of infection
- Electrical stimulation for bone healing is conditionally recommended (evidentiary support, but lack of full consensus) as an adjunct for primary bone healing of acute traumatic spinal fracture(s) in patients at high-risk for the development of pseudarthrosis. This conditional recommendation is for 6 months followed by CT scan imaging to establish effectiveness (specifically osseous union development across site of spinal fracture), and dynamic imaging failing to show translational displacement greater than 2mm and focal segmental angulation  $> 5^\circ$  exhibiting one or more high-risk criteria:
  - Acute spinal fracture as defined:
    - Acute traumatic spinal fracture (i.e., compression/burst fractures with 50% or greater loss of vertebral height, associated segmental angulation of  $5-15^\circ$ , associated stable translational deformity as demonstrated on dynamic imaging, without neurologic deficits or impending neurologic deficits) in high-risk patients for pseudarthrosis
    - Acute spondylolysis that is considered symptomatic with or without radiculopathy and without neurologic compression demonstrated on radiographic imaging in high-risk patients
    - Acute nontraumatic osteoporotic compression fractures associated with significant pain for greater than 6 weeks without neurologic deficits and without evidence for pathologic fracture (i.e., neoplasm or infection)

#### High -Risk Criteria:

- Smokers who cannot stop smoking in preparation for fusion due to the nature of the underlying condition (e.g., acute traumatic fracture)
- Exhibit one or more of the following comorbidities:

- Diabetes
- Inflammatory arthritis (e.g., rheumatoid arthritis) that has required long-term corticosteroid therapy
- Immunocompromised (e.g., undergoing chemotherapy and radiation therapy to the spine, hypogammaglobulinemia, granulocytopenia, acquired immune deficiency syndrome, chronic granulomatous disease)
- Systemic vascular disease
- Osteopenia or osteoporosis
  - Elderly or debilitated patients that surgical remedy carries predictable high risk for complications and morbidity

This coverage policy recommendation also states that even in the event of the above qualifying scenarios, the following should be specifically considered:

- Pregnancy
- Infection
- Individuals with pacemakers or defibrillators
- Children
- Individuals who will require MRI studies

## U.S. Food and Drug Administration (FDA)

This section is to be used for informational purposes only. FDA approval alone is not a basis for coverage.

The FDA regards bone growth stimulators as significant-risk (Class III) devices. Because the list of products used for bone growth stimulation is extensive, refer to the following website for more information and search by product name in the Device Name field on either the 510(k) page or on the Premarket Approvals page using Product Codes LOE (for stimulator, invasive bone growth), LOF (for stimulator, bone growth, non-invasive), or LPQ (for ultrasound bone growth stimulators): <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPMN/pmn.cfm>. (Accessed August 1, 2025)

## References

Agency for Healthcare Research and Quality (AHRQ). The role of bone growth stimulating devices and orthobiologics in healing nonunion fractures. Technology assessment program. Prepared by ECRI Evidence-based Practice Center (EPC) (Schoelles et al.). September 21, 2005.

Akhter S, Qureshi AR, Aleem I, et al. Efficacy of electrical stimulation for spinal fusion: A systematic review and meta-analysis of randomized controlled trials. *Sci Rep*. 2020 Mar 12;10(1):4568.

Aleem IS, Aleem I, Evaniew N, et al. Efficacy of electrical stimulators for bone healing: a meta-analysis of randomized sham-controlled trials. *Sci Rep*. 2016 Aug 19;6:31724.

Foley KT, Mroz TE, Arnold PM, et al. Randomized, prospective, and controlled clinical trial of pulsed electromagnetic field stimulation for cervical fusion. *Spine J*. 2008 May-Jun;8(3):436-42.

Griffin M, Bayat A. Electrical stimulation in bone healing: critical analysis by evaluating levels of evidence. *Eplasty*. 2011;11: e34. Epub 2011 Jul 26.

Haglin JM, Jain S, Eltorai AEM, Daniels AH. Bone growth stimulation: A critical analysis review. *JBJS Rev*. 2017 Aug;5(8):e8.

North American Spine Society. Electrical Stimulation for Bone Healing. Coverage Policy Recommendations. August 2016; Updated February 2025.

## Policy History/Revision Information

Date	Summary of Changes
02/01/2026	<p><b>Title Change</b></p> <ul style="list-style-type: none"> <li>● Previously titled <i>Electrical and Ultrasound Bone Growth Stimulators (for Kansas Only)</i></li> </ul> <p><b>Medical Records Documentation Used for Reviews</b></p> <ul style="list-style-type: none"> <li>● Added language to indicate:               <ul style="list-style-type: none"> <li>○ Benefit coverage for health services is determined by the federal, state, or contractual requirements, and applicable laws that may require coverage for a specific service</li> </ul> </li> </ul>

Date	Summary of Changes
	<ul style="list-style-type: none"> <li>○ Medical records documentation may be required to assess whether the member meets the clinical criteria for coverage but does not guarantee coverage of the service requested</li> <li>○ The patient's medical record must contain documentation that fully supports the medical necessity for the requested services</li> <li>○ This documentation includes but is not limited to relevant medical history, physical examination, and results of pertinent diagnostic tests or procedures</li> <li>○ Documentation supporting the medical necessity should be legible, maintained in the patient's medical record, and must be made available upon request</li> </ul> <p><b>Supporting Information</b></p> <ul style="list-style-type: none"> <li>● Updated <i>Description of Services</i>, <i>Clinical Evidence</i>, and <i>References</i> sections to reflect the most current information</li> <li>● Archived previous policy version CS037KS.01</li> </ul>

## Instructions for Use

This Medical Policy provides assistance in interpreting UnitedHealthcare standard benefit plans. When deciding coverage, the federal, state, or contractual requirements for benefit plan coverage must be referenced as the terms of the federal, state, or contractual requirements for benefit plan coverage may differ from the standard benefit plan. In the event of a conflict, the federal, state, or contractual requirements for benefit plan coverage govern. Before using this policy, please check the federal, state, or contractual requirements for benefit plan coverage. UnitedHealthcare reserves the right to modify its policies and guidelines as necessary. This Medical Policy is provided for informational purposes. It does not constitute medical advice.

UnitedHealthcare uses InterQual® for the primary medical/surgical criteria, and the American Society of Addiction Medicine (ASAM) criteria for substance use disorder (SUD) services, in administering health benefits. If InterQual® does not have applicable criteria, UnitedHealthcare may also use UnitedHealthcare Medical Policies that have been approved by the Kansas Department of Health and Environment. The UnitedHealthcare Medical Policies are intended to be used in connection with the independent professional medical judgment of a qualified health care provider and do not constitute the practice of medicine or medical advice.